



<p>ASSEMBLY INSTRUCTIONS</p>	<p>FULLY ASSEMBLED</p> <p>ASSEMBLED THERMAL ANGEL BLOOD AND IV FLUID WARMING SYSTEM</p>	<p>COMPLETE PACK</p> <p>THERMAL ANGEL ULTRA EMERGENCY PACK (TA-U)</p> <p>R E C O U R S E</p> <p>DISPOSABLE ITEM FOR PACK RELOAD</p> <p>THERMAL ANGEL DISPOSABLE PACK RELOAD KIT (TA-DEP)</p> <p>Thermal Angel (TA-200) QTY 4 Each R</p> <p>9" IV Extension Set (TA-9) QTY 4 Each E</p>
<p>CHARGING INSTRUCTIONS</p> <p>IMPORTANT</p> <p>The Thermal Angel (TA-200) will NOT operate while the Ultra Battery 1 (TA-UB1) is being charged; allow 90 min. for full recharge.</p>		<p>DISTRIBUTOR / ORDER</p> <p>Bound Tree Medical Phone: (800) 533-0523 Fax: (800) 257-5713 Web: www.BoundTree.com</p> <p>MANUFACTURER</p> <p>Estill Medical Technology Phone: (877) 354-0288 Web: www.ThermalAngel.com</p> <p>Thermal Angel Blood and IV Fluid Warming System 808-0001</p>